

PRODUCT DATASHEET

MODEL: N52BLB



Trade Name	Non-sterile Powder Free Nitrile Examination Gloves
Intended Use	A patient examination glove is a disposable device intended for medical purpose that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.
Product Conformance	EN 455 Current Revision and Customer Requirement
Material	Nitrile (Carboxylated Butadiene Acrylonitrile)
Cuff Finishing	Beaded
Color	Blue
Type	Ambidextrous
Surface finish	Finger textured

PRODUCT SPECIFICATIONS			
Dimension	Size	Palm Width (mm)	Length (mm)
	X-Small	80 ± 5mm	240min
	Small	85 ± 5mm	240min
	Medium	95 ± 5mm	240min
	Large	106 ± 5mm	240min
	Extra Large	115 ± 5mm	240min
Thickness	Single wall thickness (mm)		
	Location	Specification	
	Cuff	0.05-0.09	
	Palm	0.07-0.11	
	Finger	0.11-0.19	
Physical Properties	Beading	0.5min	
	Force at break (N)	Un-Aged	Aged
		6min	6min
Residual Powder	2.0 mg/glove Maximum		
Quality Specification * AND – Accept No Defects	Inspection Criteria		Sampling Plan/AQL
	Dimension		13 pcs (Median)
	Force at Break		13 pcs (Median)
	Water Tight Test		G1 – AQL 1.5
	Visual Inspection – Major		G1 – AQL 2.5
	Visual Inspection – Minor		G1 – AQL 4.0
	Residual Powder		n = 5 (average)
	Packaging Labeling		n = 5 (AND)
Glove Physical Counting		n = 5 (AND)	
Packing	100 Gloves per Inner Dispenser (by weight) 10 / 20 Dispensers per Outer Carton		